



Facility/Laboratory Name: _____

VOLUME CHECK FOR CONTINUOUS PIPETTING OUTFIT

Test Kit for Use _____

Pipettor ID _____

Loop drill bit fit check ^{5,6}		DATE	1	2	3	4	5	6	7	8	9	10	AVG.	Analyst initials/ID
#54	#53													
Yes	No													

1. Ten (10) volume checks are required **each day of use** for Cornwall Continuous-pipetting outfit.
2. Individual unit must be etched, identified and tagged with the average volume.
3. Tolerance permitted – 5% of volume.
4. Use a Class “A” graduated 10 ml. Dispense as used during test procedure.
5. Check loop fit over smooth part of drill bit. Loop fits over No. 54 but not a No. 53 twist drill bit.
6. Check loop fit monthly.