



Facility/Laboratory Name: \_\_\_\_\_

**CHARM ROSA READER (ROSA Reader, ROSA Pearl Reader or Charm Sciences equivalent)**

**PRIMARY CALIBRATION STRIPS**

YEAR \_\_\_\_\_ MONTH \_\_\_\_\_

SERIAL # OF PRIMARY CALIBRATION STRIPS \_\_\_\_\_

DAY	<u>LOW RANGE:</u>		<u>HIGH RANGE:</u>		ANALYST ID# OR INITIALS
	-20%	+20%	-20%	+20%	
1					
2					
3					
4					
5					
6					
7					
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					
26					
27					
28					
29					
30					
31					

1. Primary Calibration Strips need to be done day of use along with a positive and negative control.
2. Primary Calibration Strips match ROSA serial number. Calibration strips are specific to an individual reader. Do not interchange strips between different readers.